

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU112919\
 Data File : VU035921.D
 Acq On : 29 Nov 2019 14:33
 Operator : JC/SP
 Sample : K6061-03
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0XQ1

Quant Time: Nov 30 04:26:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR112119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Nov 30 03:20:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	221850	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	261406	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.84	152	80003	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	123453	4.26	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	85.20%
7) Chloroethane-d5	1.93	69	123443	4.92	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.40%
11) 1,1-Dichloroethene-d2	2.59	63	158616	2.94	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	58.80%#
20) 2-Butanone-d5	4.69	46	333257	49.69	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.38%
24) Chloroform-d	5.11	84	235171	4.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.80%
26) 1,2-Dichloroethane-d4	5.75	65	130119	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.40%
32) Benzene-d6	5.77	84	435049	4.36	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.20%
36) 1,2-Dichloropropane-d6	6.73	67	152341	4.53	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.60%
41) Toluene-d8	7.93	98	349332	3.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.00%
43) trans-1,3-Dichloropropene-	8.21	79	47854	3.66	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.20%
46) 2-Hexanone-d5	8.68	63	246046	46.81	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.62%
57) 1,1,2,2-Tetrachloroethane-	10.79	84	128967	4.56	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.20%
64) 1,2-Dichlorobenzene-d4	12.22	152	90446	5.20	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU112919\
Data File : VU035921.D
Acq On : 29 Nov 2019 14:33
Operator : JC/SP
Sample : K6061-03
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0XQ1

Quant Time: Nov 30 04:26:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR112119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Nov 30 03:20:00 2019
Response via : Initial Calibration

