

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112923\
 Data File : VU056638.D
 Acq On : 29 Nov 2023 13:43
 Operator : MD/SY
 Sample : VIBLK169
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK169

Quant Time: Nov 29 23:30:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 29 23:29:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	265629	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	266707	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	112518	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	89024	4.481	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	89.600%
7) Chloroethane-d5	1.911	69	73860	4.849	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.000%
11) 1,1-Dichloroethene-d2	2.563	65	40569	4.627	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	92.600%
20) 2-Butanone-d5	4.624	46	227159	57.338	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	114.680%
24) Chloroform-d	5.055	84	187991	4.892	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.800%
26) 1,2-Dichloroethane-d4	5.695	65	95803	5.395	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.800%
32) Benzene-d6	5.721	84	358970	4.517	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.400%
36) 1,2-Dichloropropane-d6	6.685	67	113980	4.759	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.200%
41) Toluene-d8	7.894	98	296916	4.231	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.600%
43) trans-1,3-Dichloroprop...	8.177	79	39934	4.348	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.000%
46) 2-Hexanone-d5	8.631	63	169357	46.919	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.840%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	83333	4.870	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.400%
66) 1,2-Dichlorobenzene-d4	12.190	152	104963	5.050	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112923\
Data File : VU056638.D
Acq On : 29 Nov 2023 13:43
Operator : MD/SY
Sample : VIBLK169
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK169

Quant Time: Nov 29 23:30:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 29 23:29:11 2023
Response via : Initial Calibration

