

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU113021\
 Data File : VU046028.D
 Acq On : 30 Nov 2021 12:36
 Operator : SY/MD
 Sample : M4833-19ME
 Misc : 4.27g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ESQN6ME

Quant Time: Dec 01 00:34:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Dec 01 00:33:27 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	215447	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.426	117	213009	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.822	152	101269	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	75905	42.770	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	85.540%
7) Chloroethane-d5	1.919	69	53702	39.410	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	78.820%
11) 1,1-Dichloroethene-d2	2.555	63	83579	26.233	ug/L	-0.02
Spiked Amount	50.000	Range	60 - 125	Recovery	=	52.460%#
21) 2-Butanone-d5	4.668	46	121066	85.812	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	85.810%
24) Chloroform-d	5.066	84	123733	41.749	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.500%
26) 1,2-Dichloroethane-d4	5.703	65	87488	43.986	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.980%
32) Benzene-d6	5.726	84	276379	45.269	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.540%
36) 1,2-Dichloropropane-d6	6.693	67	86098	45.504	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.000%
41) Toluene-d8	7.902	98	251598	45.261	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.520%
43) trans-1,3-Dichloroprop...	8.182	79	41563	45.504	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.000%
47) 2-Hexanone-d5	8.687	63	79152	88.229	ug/L	0.05
Spiked Amount	100.000	Range	45 - 130	Recovery	=	88.230%
56) 1,1,2,2-Tetrachloroeth...	10.774	84	118233	41.204	ug/L	0.01
Spiked Amount	50.000	Range	65 - 120	Recovery	=	82.400%
66) 1,2-Dichlorobenzene-d4	12.201	152	92162	47.256	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.520%
Target Compounds						
					Qvalue	
19) 1,1-Dichloroethane	3.870	63	3695	1.289	ug/L	94
20) cis-1,2-Dichloroethene	4.671	96	8116	4.745	ug/L	97
30) 1,1,1-Trichloroethane	5.314	97	4781	1.829	ug/L	98
34) Trichloroethene	6.542	95	13225	7.787	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU113021\
Data File : VU046028.D
Acq On : 30 Nov 2021 12:36
Operator : SY/MD
Sample : M4833-19ME
Misc : 4.27g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ESQN6ME

Quant Time: Dec 01 00:34:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Dec 01 00:33:27 2021
Response via : Initial Calibration

