

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU113021\
 Data File : VU046026.D
 Acq On : 30 Nov 2021 10:53
 Operator : SY/MD
 Sample : VU1130WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK040

Quant Time: Dec 01 00:34:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Dec 01 00:33:27 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	186349	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	188913	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	89032	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	73607	47.951	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.900%
7) Chloroethane-d5	1.919	69	58333	49.494	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.980%
11) 1,1-Dichloroethene-d2	2.571	63	100986	36.646	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.300%
21) 2-Butanone-d5	4.649	46	125095	102.513	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.510%
24) Chloroform-d	5.067	84	120130	46.863	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.720%
26) 1,2-Dichloroethane-d4	5.706	65	84869	49.332	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.660%
32) Benzene-d6	5.732	84	268996	49.680	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.360%
36) 1,2-Dichloropropane-d6	6.694	67	83778	49.925	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.860%
41) Toluene-d8	7.899	98	241116	48.908	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.820%
43) trans-1,3-Dichloroprop...	8.179	79	39140	48.317	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.640%
47) 2-Hexanone-d5	8.636	63	77565	97.488	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.490%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	120568	47.377	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.760%
66) 1,2-Dichlorobenzene-d4	12.195	152	90939	53.038	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.080%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU113021\
Data File : VU046026.D
Acq On : 30 Nov 2021 10:53
Operator : SY/MD
Sample : VU1130WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK040

Quant Time: Dec 01 00:34:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Dec 01 00:33:27 2021
Response via : Initial Calibration

