

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120122\
 Data File : VU052159.D
 Acq On : 01 Dec 2022 13:16
 Operator : JC/MD
 Sample : VSTD0.0527
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.0027

Quant Time: Dec 02 00:27:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM120122.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Dec 02 00:22:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.207	114	16631	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.349	117	14858	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.745	152	6599	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.599	65	828	0.042	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.661	65	1191	0.062	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.640	67	764	0.042	ug/L	0.00
8) Toluene-d8	7.833	98	1383	0.050	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.683	84	898	0.047	ug/L	0.00
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.599	62	1151	0.056	ug/L #	73
6) Trichloroethene	6.489	95	872	0.053	ug/L	96
9) 1,2-Dibromoethane	8.870	107	595	0.049	ug/L #	91
12) 1,2,3-Trichloropropane	10.751	75	544	0.051	ug/L	97
13) 1,2-Dibromo-3-chloropr...	12.928	75	119	0.057	ug/L #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120122\
Data File : VU052159.D
Acq On : 01 Dec 2022 13:16
Operator : JC/MD
Sample : VSTD0.0527
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.0027

Quant Time: Dec 02 00:27:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM120122.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Dec 02 00:22:18 2022
Response via : Initial Calibration

