

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120222\
 Data File : VU052171.D
 Acq On : 02 Dec 2022 11:04
 Operator : JC/MD
 Sample : VU1202WBL05
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK009

Quant Time: Dec 02 21:39:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM120122.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Dec 02 00:27:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.207	114	16523	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.361	117	14387	0.500	ug/L	0.01
11) 1,4-Dichlorobenzene-d4	11.742	152	6259	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.599	65	10687	0.549	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	110.000%
4) 1,2-Dichloroethane-d4	5.661	65	7944	0.431	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	86.000%
7) 1,2-Dichloropropane-d6	6.639	67	9505	0.550	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	110.000%
8) Toluene-d8	7.830	98	12005	0.461	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	92.000%
10) 1,1,2,2-Tetrachloroeth...	10.679	84	9258	0.501	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	100.000%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120222\
Data File : VU052171.D
Acq On : 02 Dec 2022 11:04
Operator : JC/MD
Sample : VU1202WBL05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK009

Quant Time: Dec 02 21:39:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM120122.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Dec 02 00:27:38 2022
Response via : Initial Calibration

