

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120222\
 Data File : VU052171.D
 Acq On : 02 Dec 2022 11:04
 Operator : JC/MD
 Sample : VU1202WBL05
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK009

Quant Time: Dec 02 21:39:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM120122.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Dec 02 00:27:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.207	114	16523	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.361	117	14387	0.500	ug/L	0.01
11) 1,4-Dichlorobenzene-d4	11.742	152	6259	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.599	65	10687	0.549	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery = 110.000%			
4) 1,2-Dichloroethane-d4	5.661	65	7944	0.431	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 86.000%			
7) 1,2-Dichloropropane-d6	6.639	67	9505	0.550	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery = 110.000%			
8) Toluene-d8	7.830	98	12005	0.461	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 92.000%			
10) 1,1,2,2-Tetrachloroeth...	10.679	84	9258	0.501	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery = 100.000%			

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120222\
Data File : VU052171.D
Acq On : 02 Dec 2022 11:04
Operator : JC/MD
Sample : VU1202WBL05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK009

Quant Time: Dec 02 21:39:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM120122.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Dec 02 00:27:38 2022
Response via : Initial Calibration

