

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120222\
 Data File : VU052174.D
 Acq On : 02 Dec 2022 13:38
 Operator : JC/MD
 Sample : VSTDCCC0.5EC
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.5091

Quant Time: Dec 02 21:39:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM120122.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Dec 02 00:27:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.206	114	15136	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.361	117	13557	0.500	ug/L	0.01
11) 1,4-Dichlorobenzene-d4	11.742	152	6064	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.598	65	8658	0.485	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery =	98.000%		
4) 1,2-Dichloroethane-d4	5.661	65	7265	0.430	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	86.000%		
7) 1,2-Dichloropropane-d6	6.639	67	8362	0.513	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery =	102.000%		
8) Toluene-d8	7.845	98	10664	0.434	ug/L	0.01
Spiked Amount 0.500	Range 70 - 130		Recovery =	86.000%		
10) 1,1,2,2-Tetrachloroeth...	10.679	84	8414	0.483	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery =	96.000%		
Target Compounds						
3) Vinyl chloride	1.598	62	9790	0.507	ug/L	99
6) Trichloroethene	6.489	95	7806	0.514	ug/L	100
9) 1,2-Dibromoethane	8.866	107	5809	0.509	ug/L	97
12) 1,2,3-Trichloropropane	10.748	75	4970	0.490	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.924	75	922	0.455	ug/L #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120222\
Data File : VU052174.D
Acq On : 02 Dec 2022 13:38
Operator : JC/MD
Sample : VSTDCCC0.5EC
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.5091

Quant Time: Dec 02 21:39:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM120122.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Dec 02 00:27:38 2022
Response via : Initial Calibration

