

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU120318\
 Data File : VU028436.D
 Acq On : 03 Dec 2018 13:12
 Operator : MD/SY
 Sample : PB115231ZHE#13
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB115231ZHE#13

Quant Time: Dec 04 07:19:52 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U112818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 29 01:04:49 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	295250	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	557633	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	519645	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	240296	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	241522	48.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.90%	
35) Dibromofluoromethane	4.88	113	181189	48.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.72%	
50) Toluene-d8	7.57	98	722489	48.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.44%	
62) 4-Bromofluorobenzene	10.30	95	257487	48.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.34%	
Target Compounds						
16) Acetone	2.32	43	17353	5.749	ug/l	95
20) Methylene Chloride	2.70	84	30079	7.001	ug/l	94
43) Isopropyl Acetate	5.55	43	33829	2.567	ug/l	98
95) Naphthalene	13.75	128	9580	3.282	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU120318\
Data File : VU028436.D
Acq On : 03 Dec 2018 13:12
Operator : MD/SY
Sample : PB115231ZHE#13
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB115231ZHE#13

Quant Time: Dec 04 07:19:52 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U112818W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 29 01:04:49 2018
Response via : Initial Calibration

