

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU120318\
 Data File : VU028437.D
 Acq On : 03 Dec 2018 13:36
 Operator : MD/SY
 Sample : PB115231ZHE#14
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB115231ZHE#14

Quant Time: Dec 04 07:21:21 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U112818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 29 01:04:49 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	298132	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	562536	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	525430	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	247384	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	246734	49.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.04%	
35) Dibromofluoromethane	4.88	113	187821	50.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.42%	
50) Toluene-d8	7.57	98	737898	49.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.66%	
62) 4-Bromofluorobenzene	10.31	95	265436	49.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.46%	
Target Compounds						
16) Acetone	2.32	43	16123	5.290	ug/l	89
20) Methylene Chloride	2.70	84	25306	5.821	ug/l	93
43) Isopropyl Acetate	5.55	43	32869	2.473	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU120318\
Data File : VU028437.D
Acq On : 03 Dec 2018 13:36
Operator : MD/SY
Sample : PB115231ZHE#14
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB115231ZHE#14

Quant Time: Dec 04 07:21:21 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U112818W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 29 01:04:49 2018
Response via : Initial Calibration

