

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU120318\
 Data File : VU028446.D
 Acq On : 03 Dec 2018 17:05
 Operator : MD/SY
 Sample : J5764-24
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 JC-02-113018-B

Quant Time: Dec 04 07:40:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U112818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 29 01:04:49 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	285160	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	534574	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	513140	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	247351	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	242371	50.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.72%	
35) Dibromofluoromethane	4.88	113	178527	50.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.44%	
50) Toluene-d8	7.57	98	705401	49.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.24%	
62) 4-Bromofluorobenzene	10.31	95	259878	50.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.44%	
Target Compounds						
16) Acetone	2.32	43	16978	5.824	ug/l	100
20) Methylene Chloride	2.70	84	2971255	723.590	ug/l	96
43) Isopropyl Acetate	5.55	43	34900	2.763	ug/l	98
68) m/p-Xylenes	9.38	106	8493	1.176	ug/l	78
95) Naphthalene	13.74	128	73879	6.183	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU120318\
Data File : VU028446.D
Acq On : 03 Dec 2018 17:05
Operator : MD/SY
Sample : J5764-24
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
JC-02-113018-B

Quant Time: Dec 04 07:40:05 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U112818W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 29 01:04:49 2018
Response via : Initial Calibration

