

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU120320\
 Data File : VU041456.D
 Acq On : 03 Dec 2020 17:55
 Operator : SY/MD
 Sample : I.BLK
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK04

Quant Time: Dec 04 03:04:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR120320WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Dec 04 02:22:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	57469	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	57106	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	28694	5.00	ug/L	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) Vinyl Chloride-d3	1.60	65	17177	5.44	ug/L	0.00
7) Chloroethane-d5	1.92	69	14397	5.31	ug/L	0.00
11) 1,1-Dichloroethene-d2	2.58	63	28535	3.78	ug/L	0.00
20) 2-Butanone-d5	4.65	46	56979	54.39	ug/L	0.00
24) Chloroform-d	5.08	84	40923	5.03	ug/L	0.00
26) 1,2-Dichloroethane-d4	5.72	65	27646	5.14	ug/L	0.00
32) Benzene-d6	5.74	84	65915	5.17	ug/L	0.00
36) 1,2-Dichloropropane-d6	6.70	67	18865	5.11	ug/L	0.00
41) Toluene-d8	7.91	98	60507	4.97	ug/L	0.00
43) trans-1,3-Dichloropropene-	8.19	79	11069	5.19	ug/L	0.00
46) 2-Hexanone-d5	8.64	63	43935	53.74	ug/L	0.00
57) 1,1,2,2-Tetrachloroethane-	10.75	84	19475	4.84	ug/L	0.00
64) 1,2-Dichlorobenzene-d4	12.19	152	27344	5.47	ug/L	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU120320\
Data File : VU041456.D
Acq On : 03 Dec 2020 17:55
Operator : SY/MD
Sample : I.BLK
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK04

Quant Time: Dec 04 03:04:38 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR120320WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Dec 04 02:22:08 2020
Response via : Initial Calibration

