

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120324\
 Data File : VU062077.D
 Acq On : 03 Dec 2024 09:47
 Operator : MD/SY
 Sample : VSTD0.539
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.5039

Quant Time: Dec 03 20:37:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM120324.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Dec 03 20:36:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	16418	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.419	117	23174	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	7520	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	5773	0.414	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.685	65	6211	0.552	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.687	67	6767	0.494	ug/L	0.00
8) Toluene-d8	7.887	98	17392	0.461	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.745	84	5504	0.527	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.597	62	7408	0.469	ug/L	100
6) Trichloroethene	6.544	95	6547	0.495	ug/L	100
9) 1,2-Dibromoethane	8.924	107	4363	0.519	ug/L	100
12) 1,2,3-Trichloropropane	10.814	75	4044	0.509	ug/L	100
13) 1,2-Dibromo-3-chloropr...	12.990	75	776	0.502	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120324\
Data File : VU062077.D
Acq On : 03 Dec 2024 09:47
Operator : MD/SY
Sample : VSTD0.539
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.5039

Quant Time: Dec 03 20:37:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM120324.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Dec 03 20:36:39 2024
Response via : Initial Calibration

