

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U120419\
 Data File : VU035969.D
 Acq On : 04 Dec 2019 14:06
 Operator : JC/SP
 Sample : K6056-08
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Dec 05 04:38:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR120319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Dec 05 03:54:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	288422	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	321893	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.84	152	104027	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	148901	4.53	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.60%
7) Chloroethane-d5	1.93	69	127342	4.83	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.60%
11) 1,1-Dichloroethene-d2	2.59	63	169477	3.59	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.80%
20) 2-Butanone-d5	4.70	46	292010	50.16	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.32%
24) Chloroform-d	5.11	84	225479	4.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.80%
26) 1,2-Dichloroethane-d4	5.75	65	121245	5.07	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.40%
32) Benzene-d6	5.77	84	427889	4.49	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.80%
36) 1,2-Dichloropropane-d6	6.73	67	143051	4.64	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.80%
41) Toluene-d8	7.93	98	346114	3.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.80%
43) trans-1,3-Dichloropropene-	8.21	79	53212	4.27	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.40%
46) 2-Hexanone-d5	8.68	63	212493	43.33	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.66%
57) 1,1,2,2-Tetrachloroethane-	10.78	84	120164	4.63	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.60%
64) 1,2-Dichlorobenzene-d4	12.22	152	111122	5.17	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.40%

Target Compounds					Ovalue
13) Acetone	2.69	43	9264	1.760 ug/L	99
16) Methylene chloride	3.09	84	6246	0.195 ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU120419\
Data File : VU035969.D
Acq On : 04 Dec 2019 14:06
Operator : JC/SP
Sample : K6056-08
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Dec 05 04:38:33 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR120319WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Dec 05 03:54:31 2019
Response via : Initial Calibration

