

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U120420\
 Data File : VU041475.D
 Acq On : 04 Dec 2020 17:01
 Operator : SY/MD
 Sample : L4944-08
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AQ2

Quant Time: Dec 05 06:54:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR120420WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Dec 05 05:22:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	49562	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	50382	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	25252	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	11130	4.37	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.40%
7) Chloroethane-d5	1.92	69	10803	4.84	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.80%
11) 1,1-Dichloroethene-d2	2.58	63	21865	3.55	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.00%
20) 2-Butanone-d5	4.65	46	51981	54.45	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.90%
24) Chloroform-d	5.08	84	34226	4.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.40%
26) 1,2-Dichloroethane-d4	5.72	65	24396	5.17	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.40%
32) Benzene-d6	5.74	84	53144	4.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.60%
36) 1,2-Dichloropropane-d6	6.70	67	15418	4.70	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.00%
41) Toluene-d8	7.91	98	49658	4.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.40%
43) trans-1,3-Dichloropropene-	8.19	79	9015	4.63	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.60%
46) 2-Hexanone-d5	8.64	63	37271	46.06	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.12%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	17684	4.89	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.80%
64) 1,2-Dichlorobenzene-d4	12.19	152	22563	5.22	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU120420\
Data File : VU041475.D
Acq On : 04 Dec 2020 17:01
Operator : SY/MD
Sample : L4944-08
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AQ2

Quant Time: Dec 05 06:54:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR120420WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Dec 05 05:22:35 2020
Response via : Initial Calibration

