

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U120420\
 Data File : VU041476.D
 Acq On : 04 Dec 2020 17:25
 Operator : SY/MD
 Sample : L4944-06
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AQ0

Quant Time: Dec 05 07:01:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR120420WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Dec 05 05:22:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	50352	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	47702	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	22975	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	11501	4.45	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	89.00%
7) Chloroethane-d5	1.92	69	11176	4.92	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.40%
11) 1,1-Dichloroethene-d2	2.58	63	22541	3.60	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.00%
20) 2-Butanone-d5	4.66	46	50395	51.96	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.92%
24) Chloroform-d	5.08	84	36623	5.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.60%
26) 1,2-Dichloroethane-d4	5.72	65	24978	5.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.20%
32) Benzene-d6	5.74	84	55528	5.22	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.40%
36) 1,2-Dichloropropane-d6	6.70	67	16031	5.16	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.20%
41) Toluene-d8	7.90	98	51078	4.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.20%
43) trans-1,3-Dichloropropene-	8.19	79	9647	5.24	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	104.80%
46) 2-Hexanone-d5	8.64	63	39640	51.74	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.48%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	17760	5.18	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.60%
64) 1,2-Dichlorobenzene-d4	12.19	152	23226	5.91	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	118.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU120420\
Data File : VU041476.D
Acq On : 04 Dec 2020 17:25
Operator : SY/MD
Sample : L4944-06
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AQ0

Quant Time: Dec 05 07:01:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR120420WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Dec 05 05:22:35 2020
Response via : Initial Calibration

