

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU120518\
 Data File : VU028477.D
 Acq On : 05 Dec 2018 14:01
 Operator : MD/SY
 Sample : J6208-24
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TP-6

Quant Time: Dec 06 03:08:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U112818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 29 01:04:49 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	392648	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	720300	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	668047	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	320256	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	301815	45.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.98%	
35) Dibromofluoromethane	4.88	113	92318	19.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	38.54%	
50) Toluene-d8	7.57	98	937093	48.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.84%	
62) 4-Bromofluorobenzene	10.31	95	340259	49.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.56%	
Target Compounds						
16) Acetone	2.32	43	104680	26.077	ug/l	94
20) Methylene Chloride	2.70	84	6270956	1109.142	ug/l	91
25) 2-Butanone	4.28	43	11327	1.969	ug/l	91
68) m/p-Xylenes	9.37	106	10281	1.093	ug/l	87
84) 1,2,4-Trimethylbenzene	11.15	105	34392	1.618	ug/l	99
95) Naphthalene	13.74	128	258675	11.888	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU120518\
Data File : VU028477.D
Acq On : 05 Dec 2018 14:01
Operator : MD/SY
Sample : J6208-24
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TP-6

Quant Time: Dec 06 03:08:26 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U112818W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 29 01:04:49 2018
Response via : Initial Calibration

