

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU120518\
 Data File : VU028485.D
 Acq On : 05 Dec 2018 17:07
 Operator : MD/SY
 Sample : J6208-04RE
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TP-1RE

Quant Time: Dec 06 03:33:08 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U112818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 29 01:04:49 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	355907	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	660799	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	616353	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	295160	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	284327	47.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.60%	
35) Dibromofluoromethane	4.88	113	89480	20.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	40.72%	
50) Toluene-d8	7.57	98	859591	48.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.84%	
62) 4-Bromofluorobenzene	10.30	95	309239	48.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.64%	
Target Compounds						
16) Acetone	2.32	43	134431	36.946	ug/l	98
20) Methylene Chloride	2.70	84	219575	42.774	ug/l	91
25) 2-Butanone	4.28	43	10605	2.034	ug/l #	88
95) Naphthalene	13.75	128	41980	4.429	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU120518\
Data File : VU028485.D
Acq On : 05 Dec 2018 17:07
Operator : MD/SY
Sample : J6208-04RE
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TP-1RE

Quant Time: Dec 06 03:33:08 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U112818W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 29 01:04:49 2018
Response via : Initial Calibration

