

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU120518\
 Data File : VU028489.D
 Acq On : 05 Dec 2018 18:39
 Operator : MD/SY
 Sample : J6208-24RE
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TP-6RE

Quant Time: Dec 06 03:45:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U112818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 29 01:04:49 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	341863	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.88	114	623406	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	596744	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	293326	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	271540	47.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.06%	
35) Dibromofluoromethane	4.88	113	78289	18.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	37.78%	
50) Toluene-d8	7.57	98	827405	49.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.82%	
62) 4-Bromofluorobenzene	10.30	95	305528	51.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.26%	
Target Compounds						
16) Acetone	2.32	43	94975	27.174	ug/l	99
20) Methylene Chloride	2.70	84	5599056	1137.418	ug/l	93
25) 2-Butanone	4.28	43	10760	2.148	ug/l	98
84) 1,2,4-Trimethylbenzene	11.15	105	30659	1.574	ug/l	100
95) Naphthalene	13.74	128	230156	11.629	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU120518\
Data File : VU028489.D
Acq On : 05 Dec 2018 18:39
Operator : MD/SY
Sample : J6208-24RE
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TP-6RE

Quant Time: Dec 06 03:45:57 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U112818W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 29 01:04:49 2018
Response via : Initial Calibration

