

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU120519\
 Data File : VU036004.D
 Acq On : 05 Dec 2019 19:33
 Operator : JC/SP
 Sample : K6132-07
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFQR8

Quant Time: Dec 06 02:43:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR120319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Dec 06 00:57:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	286481	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	306968	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.84	152	116026	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	130979	4.01	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.20%
7) Chloroethane-d5	1.93	69	110814	4.23	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.60%
11) 1,1-Dichloroethene-d2	2.59	63	151401	3.23	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	64.60%
20) 2-Butanone-d5	4.70	46	281276	48.64	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.28%
24) Chloroform-d	5.11	84	216477	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.60%
26) 1,2-Dichloroethane-d4	5.75	65	113983	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.00%
32) Benzene-d6	5.77	84	413552	4.55	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.00%
36) 1,2-Dichloropropane-d6	6.73	67	136375	4.64	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.80%
41) Toluene-d8	7.93	98	333445	3.98	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.60%
43) trans-1,3-Dichloropropene-	8.21	79	49472	4.16	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.20%
46) 2-Hexanone-d5	8.68	63	216849	46.37	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.74%
57) 1,1,2,2-Tetrachloroethane-	10.78	84	117505	4.75	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.00%
64) 1,2-Dichlorobenzene-d4	12.22	152	126393	5.27	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.40%

Target Compounds					Ovalue
13) Acetone	2.70	43	11938	2.283 ug/L	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU120519\
 Data File : VU036004.D
 Acq On : 05 Dec 2019 19:33
 Operator : JC/SP
 Sample : K6132-07
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampled :
 BFQR8

Quant Time: Dec 06 02:43:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR120319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Dec 06 00:57:26 2019
 Response via : Initial Calibration

