

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U120519\
 Data File : VU035992.D
 Acq On : 05 Dec 2019 14:21
 Operator : JC/SP
 Sample : K6132-10
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFQY3

Quant Time: Dec 06 01:19:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR120319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Dec 06 00:57:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	291571	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	320350	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.84	152	99829	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	137103	4.13	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.60%
7) Chloroethane-d5	1.93	69	114325	4.29	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	85.80%
11) 1,1-Dichloroethene-d2	2.59	63	149517	3.13	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.60%
20) 2-Butanone-d5	4.70	46	275716	46.85	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.70%
24) Chloroform-d	5.11	84	215492	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.40%
26) 1,2-Dichloroethane-d4	5.75	65	113198	4.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.60%
32) Benzene-d6	5.77	84	411486	4.34	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.80%
36) 1,2-Dichloropropane-d6	6.73	67	134370	4.38	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.60%
41) Toluene-d8	7.93	98	329512	3.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.40%
43) trans-1,3-Dichloropropene-	8.22	79	49629	4.00	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.00%
46) 2-Hexanone-d5	8.68	63	198589	40.69	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.38%
57) 1,1,2,2-Tetrachloroethane-	10.79	84	115140	4.46	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.20%
64) 1,2-Dichlorobenzene-d4	12.22	152	105956	5.14	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.80%


Target Compounds					Ovalue
13) Acetone	2.68	43	9097	1.710 ug/L	85


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU120519\
Data File : VU035992.D
Acq On : 05 Dec 2019 14:21
Operator : JC/SP
Sample : K6132-10
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFQY3

Quant Time: Dec 06 01:19:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR120319WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Dec 06 00:57:26 2019
Response via : Initial Calibration





#13
 Acetone
 Concen: 1.710 ug/L
 RT: 2.68 min Scan# 417
 Delta R.T. 0.00 min
 Lab File: VU035992.D
 Acq: 05 Dec 2019 14:21

Instrument :
 MSVOA_U
 ClientSampleId :
 BFQY3

Tot Ion: 43 Resp: 9097
 Ion Ratio Lower Upper
 43 100
 58 25.5 0.0 68.6

