

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120522\
 Data File : VU052179.D
 Acq On : 05 Dec 2022 12:20
 Operator : JC/MD
 Sample : MDL06
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MDL06

Quant Time: Dec 05 22:33:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM120122.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Dec 02 00:27:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.207	114	16731	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.361	117	14918	0.500	ug/L	0.01
11) 1,4-Dichlorobenzene-d4	11.742	152	6565	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.599	65	10687	0.542	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	108.000%
4) 1,2-Dichloroethane-d4	5.661	65	8205	0.439	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	88.000%
7) 1,2-Dichloropropane-d6	6.639	67	9633	0.537	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	108.000%
8) Toluene-d8	7.830	98	13836	0.512	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	102.000%
10) 1,1,2,2-Tetrachloroeth...	10.679	84	9419	0.491	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	98.000%
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.599	62	543	0.025	ug/L #	18
6) Trichloroethene	6.508	95	457	0.027	ug/L	83
9) 1,2-Dibromoethane	8.866	107	263	0.021	ug/L #	87
12) 1,2,3-Trichloropropane	10.748	75	218	0.020	ug/L #	88
13) 1,2-Dibromo-3-chloropr...	12.924	75	50	0.023	ug/L #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120522\
Data File : VU052179.D
Acq On : 05 Dec 2022 12:20
Operator : JC/MD
Sample : MDL06
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MDL06

Quant Time: Dec 05 22:33:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM120122.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Dec 02 00:27:38 2022
Response via : Initial Calibration

