

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120522\
 Data File : VU052176.D
 Acq On : 05 Dec 2022 10:47
 Operator : JC/MD
 Sample : VSTDCCC0.5
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.5092

Quant Time: Dec 05 22:33:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM120122.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Dec 02 00:27:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.206	114	17440	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.361	117	15664	0.500	ug/L	0.01
11) 1,4-Dichlorobenzene-d4	11.741	152	6732	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.598	65	11084	0.539	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	108.000%
4) 1,2-Dichloroethane-d4	5.660	65	8294	0.426	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	86.000%
7) 1,2-Dichloropropane-d6	6.639	67	9590	0.510	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	102.000%
8) Toluene-d8	7.830	98	14197	0.501	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	100.000%
10) 1,1,2,2-Tetrachloroeth...	10.679	84	9234	0.459	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	92.000%
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.598	62	10974	0.493	ug/L	94
6) Trichloroethene	6.489	95	8344	0.476	ug/L	98
9) 1,2-Dibromoethane	8.866	107	5973	0.453	ug/L	96
12) 1,2,3-Trichloropropane	10.747	75	5057	0.449	ug/L	100
13) 1,2-Dibromo-3-chloropr...	12.924	75	986	0.438	ug/L #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120522\
Data File : VU052176.D
Acq On : 05 Dec 2022 10:47
Operator : JC/MD
Sample : VSTDCCC0.5
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.5092

Quant Time: Dec 05 22:33:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM120122.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Dec 02 00:27:38 2022
Response via : Initial Calibration

