

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU120618\
 Data File : VU028505.D
 Acq On : 06 Dec 2018 16:58
 Operator : MD/SY
 Sample : J6259-22
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BP-DUP02-20181205

Quant Time: Dec 07 09:39:21 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U120618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 07 00:39:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	101489	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	195738	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	181315	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	84185	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	86445	55.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.62%	
35) Dibromofluoromethane	4.88	113	62894	51.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.12%	
50) Toluene-d8	7.57	98	253828	52.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.18%	
62) 4-Bromofluorobenzene	10.30	95	91879	50.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.90%	
Target Compounds						
44) Trichloroethene	6.19	130	5881	4.193	ug/l	95
64) Tetrachloroethene	8.23	164	1201	0.993	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU120618\
Data File : VU028505.D
Acq On : 06 Dec 2018 16:58
Operator : MD/SY
Sample : J6259-22
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BP-DUP02-20181205

Quant Time: Dec 07 09:39:21 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U120618W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 07 00:39:42 2018
Response via : Initial Calibration

