

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU120618\
 Data File : VU028513.D
 Acq On : 06 Dec 2018 20:04
 Operator : MD/SY
 Sample : J6252-12
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CPA-RB1-M16W4-11-27-18

Quant Time: Dec 07 06:54:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U120618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 07 00:39:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	95103	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	180137	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	171214	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	80104	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	82407	56.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.56%	
35) Dibromofluoromethane	4.88	113	60732	53.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.14%	
50) Toluene-d8	7.57	98	236144	52.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.32%	
62) 4-Bromofluorobenzene	10.30	95	86350	51.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.04%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU120618\
Data File : VU028513.D
Acq On : 06 Dec 2018 20:04
Operator : MD/SY
Sample : J6252-12
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CPA-RB1-M16W4-11-27-18

Quant Time: Dec 07 06:54:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U120618W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 07 00:39:42 2018
Response via : Initial Calibration

