

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U120619\
 Data File : VU036025.D
 Acq On : 06 Dec 2019 15:37
 Operator : JC/SP
 Sample : K6167-01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFQY4

Quant Time: Dec 06 23:49:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR120319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Dec 06 22:30:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	265742	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	294873	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.84	152	76806	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.62	65	110229	3.64	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	72.80%
7) Chloroethane-d5	1.93	69	130166	5.36	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.20%
11) 1,1-Dichloroethene-d2	2.59	63	133083	3.06	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.20%
20) 2-Butanone-d5	4.69	46	207598	38.70	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	77.40%
24) Chloroform-d	5.11	84	186505	4.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.60%
26) 1,2-Dichloroethane-d4	5.75	65	107742	4.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.80%
32) Benzene-d6	5.77	84	382943	4.39	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.80%
36) 1,2-Dichloropropane-d6	6.73	67	122202	4.32	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	86.40%
41) Toluene-d8	7.93	98	313067	3.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.80%
43) trans-1,3-Dichloropropene-	8.21	79	43797	3.84	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.80%
46) 2-Hexanone-d5	8.70	63	198765	44.25	ug/L	0.02
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.50%
57) 1,1,2,2-Tetrachloroethane-	10.79	84	112093	4.72	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.40%
64) 1,2-Dichlorobenzene-d4	12.22	152	84583	5.33	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.60%

Target Compounds					Ovalue
13) Acetone	2.66	43	6133	1.265 ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU120619\
Data File : VU036025.D
Acq On : 06 Dec 2019 15:37
Operator : JC/SP
Sample : K6167-01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFQY4

Quant Time: Dec 06 23:49:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR120319WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Dec 06 22:30:31 2019
Response via : Initial Calibration

