

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120621\
 Data File : VU046109.D
 Acq On : 06 Dec 2021 14:12
 Operator : SY/MD
 Sample : PB141182TB 10X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VLEB182

Quant Time: Dec 07 01:09:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 03 05:08:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	212822	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	207517	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	107011	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	73959	42.187	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	84.380%
7) Chloroethane-d5	1.915	69	54023	40.135	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	80.280%
11) 1,1-Dichloroethene-d2	2.571	63	109752	34.873	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.740%
21) 2-Butanone-d5	4.632	46	126660	90.884	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	90.880%
24) Chloroform-d	5.066	84	139544	47.665	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.320%
26) 1,2-Dichloroethane-d4	5.706	65	94365	48.029	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.060%
32) Benzene-d6	5.729	84	307741	51.740	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.480%
36) 1,2-Dichloropropane-d6	6.693	67	94271	51.142	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	102.280%
41) Toluene-d8	7.899	98	282246	52.119	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.240%
43) trans-1,3-Dichloroprop...	8.179	79	44707	50.242	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	100.480%
47) 2-Hexanone-d5	8.632	63	92746	106.118	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	106.120%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	133745	47.843	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.680%
66) 1,2-Dichlorobenzene-d4	12.195	152	111083	53.902	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120621\
Data File : VU046109.D
Acq On : 06 Dec 2021 14:12
Operator : SY/MD
Sample : PB141182TB 10X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VLEB182

Quant Time: Dec 07 01:09:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 03 05:08:36 2021
Response via : Initial Calibration

