

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120621\
 Data File : VU046104.D
 Acq On : 06 Dec 2021 11:41
 Operator : SY/MD
 Sample : VU1206MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK049

Quant Time: Dec 06 14:18:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 03 05:08:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	263169	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	255791	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.815	152	134917	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	73982	34.127	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	68.260%
7) Chloroethane-d5	1.919	69	55236	33.186	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	66.380%#
11) 1,1-Dichloroethene-d2	2.571	63	109567	28.154	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	56.300%#
21) 2-Butanone-d5	4.636	46	124726	72.375	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	72.370%
24) Chloroform-d	5.066	84	137020	37.849	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	75.700%
26) 1,2-Dichloroethane-d4	5.706	65	92232	37.962	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	75.920%
32) Benzene-d6	5.732	84	302088	41.205	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.400%
36) 1,2-Dichloropropane-d6	6.693	67	91172	40.126	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	80.260%
41) Toluene-d8	7.902	98	278090	41.660	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.320%
43) trans-1,3-Dichloroprop...	8.182	79	44773	40.820	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.640%
47) 2-Hexanone-d5	8.635	63	87066	80.819	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	80.820%
56) 1,1,2,2-Tetrachloroeth...	10.761	84	133215	38.660	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	77.320%
66) 1,2-Dichlorobenzene-d4	12.195	152	112249	43.202	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120621\
Data File : VU046104.D
Acq On : 06 Dec 2021 11:41
Operator : SY/MD
Sample : VU1206MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK049

Quant Time: Dec 06 14:18:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 03 05:08:36 2021
Response via : Initial Calibration

