

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120623\
 Data File : VU056665.D
 Acq On : 06 Dec 2023 15:44
 Operator : MD/SY
 Sample : 05642-08
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0G40

Quant Time: Dec 07 04:54:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Dec 07 04:51:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	266106	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	266478	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	119955	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	66376	3.335	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	66.600%
7) Chloroethane-d5	1.911	69	59645	3.908	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.200%
11) 1,1-Dichloroethene-d2	2.563	65	32690	3.722	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.400%
20) 2-Butanone-d5	4.628	46	226335	57.028	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	114.060%
24) Chloroform-d	5.055	84	165674	4.303	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.000%
26) 1,2-Dichloroethane-d4	5.698	65	83273	4.681	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.600%
32) Benzene-d6	5.721	84	304751	3.838	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	76.800%
36) 1,2-Dichloropropane-d6	6.685	67	102134	4.268	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.400%
41) Toluene-d8	7.894	98	244762	3.491	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	69.800%#
43) trans-1,3-Dichloroprop...	8.177	79	35139	3.829	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.600%
46) 2-Hexanone-d5	8.631	63	160573	44.524	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.040%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	77693	4.544	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.800%
66) 1,2-Dichlorobenzene-d4	12.190	152	89751	4.050	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	81.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120623\
Data File : VU056665.D
Acq On : 06 Dec 2023 15:44
Operator : MD/SY
Sample : 05642-08
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0G40

Quant Time: Dec 07 04:54:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Dec 07 04:51:32 2023
Response via : Initial Calibration

