

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU120718\
 Data File : VU028528.D
 Acq On : 07 Dec 2018 13:13
 Operator : MD/SY
 Sample : VU1207MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK51

Quant Time: Dec 11 01:02:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM120718WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 11 00:46:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	152658	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	142885	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	66226	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	70439	49.41	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	98.82%
7) Chloroethane-d5	1.68	69	54550	51.04	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	102.08%
11) 1,1-Dichloroethene-d2	2.27	63	96085	38.22	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.44%
21) 2-Butanone-d5	4.17	46	122788	107.24	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	107.24%
24) Chloroform-d	4.65	84	110170	48.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.64%
26) 1,2-Dichloroethane-d4	5.31	65	76101	51.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.32%
32) Benzene-d6	5.34	84	235312	52.67	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.34%
36) 1,2-Dichloropropane-d6	6.33	67	84850	52.88	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	105.76%
41) Toluene-d8	7.56	98	206480	50.69	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.38%
43) trans-1,3-Dichloropropene-	7.85	79	39014	53.17	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	106.34%
47) 2-Hexanone-d5	8.31	63	86097	110.81	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	110.81%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	107297	51.48	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	102.96%
64) 1,2-Dichlorobenzene-d4	11.86	152	70729	52.64	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.28%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU120718\
Data File : VU028528.D
Acq On : 07 Dec 2018 13:13
Operator : MD/SY
Sample : VU1207MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK51

Quant Time: Dec 11 01:02:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM120718WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 11 00:46:45 2018
Response via : Initial Calibration

