

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120721\
 Data File : VU046131.D
 Acq On : 07 Dec 2021 11:38
 Operator : SY/MD
 Sample : VU1207WBL02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 30 Sample Multiplier: 1

Quant Time: Dec 08 04:52:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 03 05:08:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	233127	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	229036	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	120396	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	91348	47.568	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.140%
7) Chloroethane-d5	1.919	69	68426	46.408	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.820%
11) 1,1-Dichloroethene-d2	2.572	63	126404	36.666	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.340%
21) 2-Butanone-d5	4.639	46	144636	94.743	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	94.740%
24) Chloroform-d	5.067	84	151488	47.238	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.480%
26) 1,2-Dichloroethane-d4	5.703	65	106124	49.309	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.620%
32) Benzene-d6	5.732	84	344096	52.417	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.840%
36) 1,2-Dichloropropane-d6	6.694	67	102945	50.600	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.200%
41) Toluene-d8	7.899	98	320096	53.554	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.100%
43) trans-1,3-Dichloroprop...	8.179	79	50216	51.131	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	102.260%
47) 2-Hexanone-d5	8.636	63	105096	108.951	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	108.950%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	153002	49.590	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.180%
66) 1,2-Dichlorobenzene-d4	12.195	152	127915	55.169	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	110.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120721\
Data File : VU046131.D
Acq On : 07 Dec 2021 11:38
Operator : SY/MD
Sample : VU1207WBL02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 30 Sample Multiplier: 1

Quant Time: Dec 08 04:52:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 03 05:08:36 2021
Response via : Initial Calibration

