

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120721\
 Data File : VU046158.D
 Acq On : 07 Dec 2021 22:33
 Operator : SY/MD
 Sample : M4960-06 10X
 Misc : 6.25g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 42 Sample Multiplier: 1

Quant Time: Dec 08 08:12:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 03 05:08:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	243177	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	236465	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	127750	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	80857	40.365	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	80.720%
7) Chloroethane-d5	1.919	69	58656	38.137	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	76.280%
11) 1,1-Dichloroethene-d2	2.572	63	109406	30.424	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	60.840%
21) 2-Butanone-d5	4.639	46	130242	81.789	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	81.790%
24) Chloroform-d	5.067	84	131953	39.446	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	78.900%
26) 1,2-Dichloroethane-d4	5.703	65	92422	41.168	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.340%
32) Benzene-d6	5.729	84	306621	45.241	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.480%
36) 1,2-Dichloropropane-d6	6.694	67	93638	44.580	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	89.160%
41) Toluene-d8	7.899	98	279467	45.288	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.580%
43) trans-1,3-Dichloroprop...	8.179	79	43187	42.592	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.180%
47) 2-Hexanone-d5	8.636	63	95170	95.561	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	95.560%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	138306	43.418	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.840%
66) 1,2-Dichlorobenzene-d4	12.195	152	115933	47.123	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120721\
Data File : VU046158.D
Acq On : 07 Dec 2021 22:33
Operator : SY/MD
Sample : M4960-06 10X
Misc : 6.25g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 42 Sample Multiplier: 1

Quant Time: Dec 08 08:12:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 03 05:08:36 2021
Response via : Initial Calibration

