

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120721\
 Data File : VU046159.D
 Acq On : 07 Dec 2021 22:55
 Operator : SY/MD
 Sample : M4960-05 10X
 Misc : 6.57g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGKS0

Quant Time: Dec 08 08:12:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 03 05:08:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	245988	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	239821	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	129542	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	83276	41.097	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	82.200%
7) Chloroethane-d5	1.919	69	58921	37.872	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	75.740%
11) 1,1-Dichloroethene-d2	2.571	63	113608	31.231	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	62.460%
21) 2-Butanone-d5	4.639	46	130362	80.929	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	80.930%
24) Chloroform-d	5.066	84	136747	40.412	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.820%
26) 1,2-Dichloroethane-d4	5.703	65	93867	41.334	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.660%
32) Benzene-d6	5.729	84	312882	45.519	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.040%
36) 1,2-Dichloropropane-d6	6.693	67	94526	44.373	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	88.740%
41) Toluene-d8	7.899	98	288433	46.087	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.180%
43) trans-1,3-Dichloroprop...	8.179	79	44712	43.479	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.960%
47) 2-Hexanone-d5	8.636	63	97121	96.155	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.160%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	139284	43.113	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.220%
66) 1,2-Dichlorobenzene-d4	12.195	152	118185	47.374	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.740%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120721\
Data File : VU046159.D
Acq On : 07 Dec 2021 22:55
Operator : SY/MD
Sample : M4960-05 10X
Misc : 6.57g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BGKS0

Quant Time: Dec 08 08:12:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 03 05:08:36 2021
Response via : Initial Calibration

