

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120722\
 Data File : VU052203.D
 Acq On : 07 Dec 2022 17:57
 Operator : JC/MD
 Sample : N5930-07
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EXZ92

Quant Time: Dec 08 02:02:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Dec 08 01:59:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	199689	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	208448	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	79983	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	88085	4.535	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.600%
7) Chloroethane-d5	1.916	69	80098	5.309	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	106.200%
11) 1,1-Dichloroethene-d2	2.568	65	32547	4.675	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	93.400%
20) 2-Butanone-d5	4.633	46	204735	58.150	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	116.300%
24) Chloroform-d	5.063	84	155472	5.448	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.000%
26) 1,2-Dichloroethane-d4	5.700	65	79899	5.543	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.800%
32) Benzene-d6	5.726	84	305805	4.942	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.800%
36) 1,2-Dichloropropane-d6	6.687	67	101883	5.320	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.400%
41) Toluene-d8	7.896	98	225893	4.034	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.600%
43) trans-1,3-Dichloroprop...	8.179	79	32997	4.728	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.600%
46) 2-Hexanone-d5	8.632	63	143153	48.401	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.800%
56) 1,1,2,2-Tetrachloroeth...	10.755	84	89216	5.373	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	107.400%
66) 1,2-Dichlorobenzene-d4	12.192	152	86225	5.631	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.600%
Target Compounds					Qvalue	
13) Acetone	2.655	43	6020	2.518	ug/L	67
25) Chloroform	5.086	83	11843	0.418	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120722\
Data File : VU052203.D
Acq On : 07 Dec 2022 17:57
Operator : JC/MD
Sample : N5930-07
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EXZ92

Quant Time: Dec 08 02:02:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Dec 08 01:59:42 2022
Response via : Initial Calibration

