

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120722\
 Data File : VU052209.D
 Acq On : 07 Dec 2022 20:17
 Operator : JC/MD
 Sample : N5916-05
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F5N23

Quant Time: Dec 08 02:03:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Dec 08 01:59:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	204302	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	207357	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	78761	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	83219	4.188	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.800%
7) Chloroethane-d5	1.916	69	75912	4.918	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.400%
11) 1,1-Dichloroethene-d2	2.568	65	31558	4.430	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	88.600%
20) 2-Butanone-d5	4.632	46	211945	58.838	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	117.680%
24) Chloroform-d	5.060	84	152722	5.230	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.600%
26) 1,2-Dichloroethane-d4	5.700	65	77567	5.259	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.200%
32) Benzene-d6	5.726	84	297160	4.827	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.600%
36) 1,2-Dichloropropane-d6	6.687	67	101768	5.342	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.800%
41) Toluene-d8	7.896	98	226225	4.062	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.200%
43) trans-1,3-Dichloroprop...	8.179	79	31559	4.546	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.000%
46) 2-Hexanone-d5	8.632	63	152002	51.663	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.320%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	88006	5.328	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.600%
66) 1,2-Dichlorobenzene-d4	12.192	152	82938	5.500	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.000%
Target Compounds						Qvalue
5) Vinyl chloride	1.604	62	20095	0.904	ug/L	89
10) 1,1,2-Trichloro-1,2,2-...	2.584	101	64839	4.339	ug/L	99
12) 1,1-Dichloroethene	2.584	96	8215	0.554	ug/L #	1
17) Methyl tert-butyl Ether	3.366	73	8467	0.255	ug/L	99
18) trans-1,2-Dichloroethene	3.356	96	7719	0.486	ug/L	90
19) 1,1-Dichloroethane	3.867	63	17944	0.627	ug/L	97
22) cis-1,2-Dichloroethene	4.665	96	243418	14.599	ug/L	98
27) 1,2-Dichloroethane	5.793	62	18022	1.014	ug/L	97
33) Benzene	5.777	78	8959	0.133	ug/L	100
34) Trichloroethene	6.539	95	241299	14.336	ug/L	96
37) 1,2-Dichloropropane	6.787	63	6853	0.392	ug/L #	94
47) Tetrachloroethene	8.552	164	10501	0.814	ug/L	97
51) Chlorobenzene	9.443	112	56173	1.269	ug/L	98
64) 1,3-Dichlorobenzene	11.745	146	8313	0.329	ug/L	95
65) 1,4-Dichlorobenzene	11.835	146	14720	0.584	ug/L	99
67) 1,2-Dichlorobenzene	12.211	146	21920	0.901	ug/L	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed


Instrument :
MSVOA_U
ClientSampleId :
F5N23

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120722\
Data File : VU052209.D
Acq On : 07 Dec 2022 20:17
Operator : JC/MD
Sample : N5916-05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F5N23

Quant Time: Dec 08 02:03:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Dec 08 01:59:42 2022
Response via : Initial Calibration









#5
 Vinyl chloride
 Concen: 0.904 ug/L
 RT: 1.604 min Scan# 81
 Delta R.T. -0.000 min
 Lab File: VU052209.D
 Acq: 07 Dec 2022 20:17




Instrument :
 MSVOA_U
 ClientSampleId :
 F5N23


Tgt Ion: 62 Resp: 20095
 Ion Ratio Lower Upper
 62 100
 64 39.7 23.5 43.7



#10
 1,1,2-Trichloro-1,2,2-trifluoroethane
 Concen: 4.339 ug/L
 RT: 2.584 min Scan# 387
 Delta R.T. 0.006 min
 Lab File: VU052209.D
 Acq: 07 Dec 2022 20:17

Tgt Ion: 101 Resp: 64839
 Ion Ratio Lower Upper
 101 100
 85 43.2 33.8 50.6
 151 73.1 58.6 88.0









#12
1,1-Dichloroethene
Concen: 0.554 ug/L
RT: 2.584 min Scan# 31
Delta R.T. 0.006 min
Lab File: VU052209.D
Acq: 07 Dec 2022 20:17




Instrument :
MSVOA_U
ClientSampleId :
F5N23


Tgt Ion: 96 Resp: 8215
Ion Ratio Lower Upper
96 100
61 197.5 114.8 213.2
63 495.5 93.7 174.1#



#17
Methyl tert-butyl Ether
Concen: 0.255 ug/L
RT: 3.366 min Scan# 630
Delta R.T. 0.003 min
Lab File: VU052209.D
Acq: 07 Dec 2022 20:17

Tgt Ion: 73 Resp: 8467
Ion Ratio Lower Upper
73 100
43 15.9 13.2 19.8
57 22.1 17.8 26.8









#18
trans-1,2-Dichloroethene
Concen: 0.486 ug/L
RT: 3.356 min Scan# 61
Delta R.T. 0.003 min
Lab File: VU052209.D
Acq: 07 Dec 2022 20:17




Instrument :
MSVOA_U
ClientSampleId :
F5N23


Tgt Ion: 96 Resp: 7719
Ion Ratio Lower Upper
96 100
61 124.9 97.9 181.7
98 65.9 43.8 81.3



#19
1,1-Dichloroethane
Concen: 0.627 ug/L
RT: 3.867 min Scan# 786
Delta R.T. -0.000 min
Lab File: VU052209.D
Acq: 07 Dec 2022 20:17

Tgt Ion: 63 Resp: 17944
Ion Ratio Lower Upper
63 100
65 30.5 22.2 41.2
83 11.7 9.2 17.0









#22
 cis-1,2-Dichloroethene
 Concen: 14.599 ug/L
 RT: 4.665 min Scan# 1034
 Delta R.T. -0.003 min
 Lab File: VU052209.D
 Acq: 07 Dec 2022 20:17




Instrument :
 MSVOA_U
 ClientSampleId :
 F5N23


Tgt Ion: 96 Resp: 243418
 Ion Ratio Lower Upper
 96 100
 61 120.5 85.1 158.1
 98 63.5 42.8 79.6



#27
 1,2-Dichloroethane
 Concen: 1.014 ug/L
 RT: 5.793 min Scan# 1385
 Delta R.T. -0.000 min
 Lab File: VU052209.D
 Acq: 07 Dec 2022 20:17

Tgt Ion: 62 Resp: 18022
 Ion Ratio Lower Upper
 62 100
 98 11.5 8.4 12.6









#33
Benzene
Concen: 0.133 ug/L
RT: 5.777 min Scan# 11
Delta R.T. 0.003 min
Lab File: VU052209.D
Acq: 07 Dec 2022 20:17

Instrument :
MSVOA_U
ClientSampleId :
F5N23




Tgt Ion: 78 Resp: 8959




#34
Trichloroethene
Concen: 14.336 ug/L
RT: 6.539 min Scan# 1617
Delta R.T. -0.003 min
Lab File: VU052209.D
Acq: 07 Dec 2022 20:17

Tgt Ion: 95 Resp: 241299

Ion	Ratio	Lower	Upper
95	100		
97	63.7	43.7	81.1
132	97.3	64.3	119.3
130	99.0	67.3	124.9









#37
1,2-Dichloropropane
Concen: 0.392 ug/L
RT: 6.787 min Scan# 1
Delta R.T. -0.003 min
Lab File: VU052209.D
Acq: 07 Dec 2022 20:17




Instrument :
MSVOA_U
ClientSampleId :
F5N23


Tgt Ion: 63 Resp: 6853
Ion Ratio Lower Upper
63 100
112 2.2 3.3 4.9#



#47
Tetrachloroethene
Concen: 0.814 ug/L
RT: 8.552 min Scan# 2243
Delta R.T. -0.000 min
Lab File: VU052209.D
Acq: 07 Dec 2022 20:17

Tgt Ion: 164 Resp: 10501
Ion Ratio Lower Upper
164 100
129 92.3 67.4 125.2
131 98.1 67.1 124.7
166 132.5 89.7 166.5









#51
Chlorobenzene
Concen: 1.269 ug/L
RT: 9.443 min Scan# 21
Delta R.T. -0.003 min
Lab File: VU052209.D
Acq: 07 Dec 2022 20:17




Instrument :
MSVOA_U
ClientSampleId :
F5N23


Tgt Ion:112 Resp: 56173
Ion Ratio Lower Upper
112 100
114 34.0 22.2 41.2
77 59.4 46.6 70.0



#64
1,3-Dichlorobenzene
Concen: 0.329 ug/L
RT: 11.745 min Scan# 3236
Delta R.T. -0.000 min
Lab File: VU052209.D
Acq: 07 Dec 2022 20:17

Tgt Ion:146 Resp: 8313
Ion Ratio Lower Upper
146 100
111 43.5 28.6 53.2
148 60.3 45.4 84.2









#65
1,4-Dichlorobenzene
Concen: 0.584 ug/L
RT: 11.835 min Scan# 31
Delta R.T. -0.000 min
Lab File: VU052209.D
Acq: 07 Dec 2022 20:17

Instrument :
MSVOA_U
ClientSampleId :
F5N23

Tgt Ion:146 Resp: 14720
Ion Ratio Lower Upper
146 100
111 38.5 26.9 50.1
148 62.7 45.1 83.7



#67
1,2-Dichlorobenzene
Concen: 0.901 ug/L
RT: 12.211 min Scan# 3381
Delta R.T. -0.000 min
Lab File: VU052209.D
Acq: 07 Dec 2022 20:17

Tgt Ion:146 Resp: 21920
Ion Ratio Lower Upper
146 100
111 40.1 32.7 49.1
148 60.9 50.6 75.8

