

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120722\
 Data File : VU052210.D
 Acq On : 07 Dec 2022 20:42
 Operator : JC/MD
 Sample : N5916-06
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F5N24

Quant Time: Dec 08 02:03:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Dec 08 01:59:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	205359	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	213602	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	77840	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	86816	4.346	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.000%
7) Chloroethane-d5	1.916	69	74501	4.802	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.000%
11) 1,1-Dichloroethene-d2	2.569	65	32809	4.582	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	91.600%
20) 2-Butanone-d5	4.639	46	218644	60.386	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	120.780%
24) Chloroform-d	5.060	84	153832	5.241	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.800%
26) 1,2-Dichloroethane-d4	5.700	65	80266	5.414	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.200%
32) Benzene-d6	5.726	84	300791	4.743	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.800%
36) 1,2-Dichloropropane-d6	6.691	67	101591	5.177	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.600%
41) Toluene-d8	7.896	98	225881	3.937	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.800%
43) trans-1,3-Dichloroprop...	8.179	79	31915	4.463	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.200%
46) 2-Hexanone-d5	8.636	63	154238	50.891	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.780%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	91793	5.395	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	107.800%
66) 1,2-Dichlorobenzene-d4	12.192	152	81739	5.485	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120722\
Data File : VU052210.D
Acq On : 07 Dec 2022 20:42
Operator : JC/MD
Sample : N5916-06
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F5N24

Quant Time: Dec 08 02:03:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Dec 08 01:59:42 2022
Response via : Initial Calibration

