

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120723\
 Data File : VU056684.D
 Acq On : 07 Dec 2023 16:28
 Operator : MD/SY
 Sample : 05687-01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0G66

Quant Time: Dec 08 02:04:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Dec 07 04:51:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	226693	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	234219	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	103182	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	59033	3.482	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	69.600%
7) Chloroethane-d5	1.911	69	48737	3.749	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	75.000%
11) 1,1-Dichloroethene-d2	2.563	65	29381	3.926	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.600%
20) 2-Butanone-d5	4.621	46	215263	63.668	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	127.340%
24) Chloroform-d	5.055	84	156339	4.767	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.400%
26) 1,2-Dichloroethane-d4	5.695	65	76041	5.018	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.400%
32) Benzene-d6	5.721	84	274516	3.934	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.600%
36) 1,2-Dichloropropane-d6	6.685	67	92367	4.392	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.800%
41) Toluene-d8	7.894	98	221871	3.601	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	72.000%
43) trans-1,3-Dichloroprop...	8.177	79	32112	3.981	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.600%
46) 2-Hexanone-d5	8.631	63	151807	47.891	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.780%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	74887	4.983	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	82835	4.346	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	87.000%
Target Compounds						
					Qvalue	
34) Trichloroethene	6.541	95	8818	0.484	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120723\
Data File : VU056684.D
Acq On : 07 Dec 2023 16:28
Operator : MD/SY
Sample : 05687-01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0G66

Quant Time: Dec 08 02:04:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Dec 07 04:51:32 2023
Response via : Initial Calibration

