

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120723\
 Data File : VU056691.D
 Acq On : 07 Dec 2023 19:18
 Operator : MD/SY
 Sample : 05687-08
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0G87

Quant Time: Dec 08 02:05:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Dec 07 04:51:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	229832	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	232297	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	102624	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	55580	3.233	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	64.600%
7) Chloroethane-d5	1.911	69	46004	3.490	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	69.800%
11) 1,1-Dichloroethene-d2	2.563	65	26853	3.540	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.800%
20) 2-Butanone-d5	4.618	46	207968	60.670	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	121.340%
24) Chloroform-d	5.055	84	145756	4.384	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.600%
26) 1,2-Dichloroethane-d4	5.695	65	73509	4.784	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.600%
32) Benzene-d6	5.721	84	263107	3.802	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	76.000%
36) 1,2-Dichloropropane-d6	6.682	67	89154	4.274	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.400%
41) Toluene-d8	7.894	98	211227	3.456	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	69.200%#
43) trans-1,3-Dichloroprop...	8.177	79	31496	3.937	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.800%
46) 2-Hexanone-d5	8.627	63	149222	47.465	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.920%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	71199	4.777	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	80531	4.248	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	85.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120723\
Data File : VU056691.D
Acq On : 07 Dec 2023 19:18
Operator : MD/SY
Sample : 05687-08
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0G87

Quant Time: Dec 08 02:05:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Dec 07 04:51:32 2023
Response via : Initial Calibration

