

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU120820\
 Data File : VU041498.D
 Acq On : 08 Dec 2020 12:37
 Operator : SY/MD
 Sample : VSTD00231
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD002201

Quant Time: Dec 09 00:18:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SFAMUSIM120820.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Dec 09 00:17:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.23	114	7932	0.50	ug/L	0.00
5) Chlorobenzene-d5	9.36	117	6557	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.74	152	4142	0.50	ug/L	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
2) Vinyl Chloride-d3	1.61	65	20770	2.62	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.68	65	18387	1.70	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.66	67	18034	2.29	ug/L	0.00
8) Toluene-d8	7.85	98	30777	2.10	ug/L	0.00
10) 1,1,2,2-Tetrachloroethane-	10.68	84	17565	2.17	ug/L	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
3) Vinyl chloride	1.61	62	21088	1.883	ug/L	98
6) Trichloroethene	6.51	95	17360	2.107	ug/L	98
9) 1,2-Dibromoethane	8.87	107	12169	2.067	ug/L #	99
12) 1,2,3-Trichloropropane	10.75	75	12030	1.814	ug/L	98
13) 1,2-Dibromo-3-chloropropan	12.91	75	2494	1.897	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU120820\
Data File : VU041498.D
Acq On : 08 Dec 2020 12:37
Operator : SY/MD
Sample : VSTD00231
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD002201

Quant Time: Dec 09 00:18:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SFAMUSIM120820.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Dec 09 00:17:39 2020
Response via : Initial Calibration

