

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120821\
 Data File : VU046164.D
 Acq On : 08 Dec 2021 10:33
 Operator : SY/MD
 Sample : VU1208MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK047

Quant Time: Dec 09 03:18:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 03 05:08:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	207897	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	203869	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.816	152	105917	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	76018	44.389	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.780%
7) Chloroethane-d5	1.919	69	57766	43.932	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	87.860%
11) 1,1-Dichloroethene-d2	2.572	63	105904	34.448	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.900%
21) 2-Butanone-d5	4.642	46	134676	98.925	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	98.930%
24) Chloroform-d	5.067	84	133758	46.771	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.540%
26) 1,2-Dichloroethane-d4	5.707	65	92524	48.207	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.420%
32) Benzene-d6	5.732	84	294143	50.339	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.680%
36) 1,2-Dichloropropane-d6	6.694	67	89614	49.485	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	98.980%
41) Toluene-d8	7.903	98	270474	50.839	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.680%
43) trans-1,3-Dichloroprop...	8.182	79	42475	48.588	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.180%
47) 2-Hexanone-d5	8.636	63	90746	105.688	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	105.690%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	135351	49.284	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.560%
66) 1,2-Dichlorobenzene-d4	12.195	152	106735	52.327	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.660%
Target Compounds						
72) 1,2,3-Trichlorobenzene	14.330	180	1769	0.820	ug/L	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120821\
Data File : VU046164.D
Acq On : 08 Dec 2021 10:33
Operator : SY/MD
Sample : VU1208MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK047

Quant Time: Dec 09 03:18:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 03 05:08:36 2021
Response via : Initial Calibration

