

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120822\
 Data File : VU052252.D
 Acq On : 08 Dec 2022 16:34
 Operator : JC/MD
 Sample : N5960-01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 COL24

Quant Time: Dec 08 22:32:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Dec 08 04:19:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	243075	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	256596	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	108863	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	71404	3.020	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	60.400%
7) Chloroethane-d5	1.916	69	69206	3.768	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	75.400%
11) 1,1-Dichloroethene-d2	2.568	65	29325	3.460	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.200%
20) 2-Butanone-d5	4.639	46	238110	55.558	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	111.120%
24) Chloroform-d	5.060	84	173302	4.988	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.800%
26) 1,2-Dichloroethane-d4	5.700	65	79221	4.515	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.200%
32) Benzene-d6	5.726	84	300471	3.944	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.800%
36) 1,2-Dichloropropane-d6	6.687	67	103169	4.376	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.600%
41) Toluene-d8	7.896	98	224950	3.264	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	65.200%#
43) trans-1,3-Dichloroprop...	8.176	79	31117	3.622	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	72.400%
46) 2-Hexanone-d5	8.633	63	164058	45.061	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.120%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	96622	4.727	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.600%
66) 1,2-Dichlorobenzene-d4	12.189	152	96162	4.614	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.200%
Target Compounds						Qvalue
3) Chloromethane	1.523	50	6165	0.224	ug/L	99
25) Chloroform	5.086	83	446533	12.935	ug/L	98
38) Bromodichloromethane	7.102	83	50817	1.928	ug/L	95
42) Toluene	7.967	91	20288	0.236	ug/L	99
67) 1,2-Dichlorobenzene	12.205	146	6348	0.189	ug/L #	83


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120822\
Data File : VU052252.D
Acq On : 08 Dec 2022 16:34
Operator : JC/MD
Sample : N5960-01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
COL24

Quant Time: Dec 08 22:32:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Dec 08 04:19:43 2022
Response via : Initial Calibration









#3
Chloromethane
Concen: 0.224 ug/L
RT: 1.523 min Scan# 51
Delta R.T. 0.003 min
Lab File: VU052252.D
Acq: 08 Dec 2022 16:34




Instrument :
MSVOA_U
ClientSampleId :
COL24


Tgt Ion: 50 Resp: 6165
Ion Ratio Lower Upper
50 100
52 33.8 23.4 43.4



#25
Chloroform
Concen: 12.935 ug/L
RT: 5.086 min Scan# 1165
Delta R.T. 0.000 min
Lab File: VU052252.D
Acq: 08 Dec 2022 16:34

Tgt Ion: 83 Resp: 446533
Ion Ratio Lower Upper
83 100
85 64.4 43.8 81.4





#38

Bromodichloromethane

Concen: 1.928 ug/L

RT: 7.102 min Scan# 1

Delta R.T. 0.000 min

Lab File: VU052252.D

Acq: 08 Dec 2022 16:34

Instrument :

MSVOA_U

ClientSampleId :

COL24





Tgt Ion: 83 Resp: 50817

Ion Ratio Lower Upper

83 100

85 67.0 43.5 80.9

127 8.5 6.8 10.2



#42

Toluene

Concen: 0.236 ug/L

RT: 7.967 min Scan# 2061

Delta R.T. 0.000 min

Lab File: VU052252.D




Acq: 08 Dec 2022 16:34


Tgt Ion: 91 Resp: 20288

Ion Ratio Lower Upper

91 100

92 58.8 40.7 75.5





#67
1,2-Dichlorobenzene
Concen: 0.189 ug/L
RT: 12.205 min Scan# 31
Delta R.T. -0.003 min
Lab File: VU052252.D
Acq: 08 Dec 2022 16:34

Instrument :
MSVOA_U
ClientSampleId :
COL24

Tgt Ion:146 Resp: 6348
Ion Ratio Lower Upper
146 100
111 56.0 32.7 49.1#
148 73.2 50.6 75.8

