

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120823\
 Data File : VU056711.D
 Acq On : 08 Dec 2023 19:53
 Operator : MD/SY
 Sample : 05728-01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0GA8

Quant Time: Dec 08 21:00:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 08 20:56:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	231549	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	241988	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	99581	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	97573	5.634	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	112.600%
7) Chloroethane-d5	1.907	69	74441	5.606	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	112.200%
11) 1,1-Dichloroethene-d2	2.563	65	41464	5.425	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	108.400%
20) 2-Butanone-d5	4.618	46	186104	53.889	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.780%
24) Chloroform-d	5.055	84	175860	5.250	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.000%
26) 1,2-Dichloroethane-d4	5.695	65	87004	5.621	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	112.400%
32) Benzene-d6	5.721	84	342132	4.745	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.000%
36) 1,2-Dichloropropane-d6	6.685	67	108271	4.983	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.600%
41) Toluene-d8	7.894	98	287692	4.519	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.400%
43) trans-1,3-Dichloroprop...	8.177	79	39497	4.739	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.800%
46) 2-Hexanone-d5	8.627	63	138813	42.386	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.780%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	75524	4.864	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.200%
66) 1,2-Dichlorobenzene-d4	12.190	152	97279	5.288	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120823\
Data File : VU056711.D
Acq On : 08 Dec 2023 19:53
Operator : MD/SY
Sample : 05728-01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0GA8

Quant Time: Dec 08 21:00:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Dec 08 20:56:24 2023
Response via : Initial Calibration

