

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU120919\
 Data File : VU036053.D
 Acq On : 09 Dec 2019 17:38
 Operator : JC/MD
 Sample : K6168-11
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFE44

Quant Time: Dec 10 02:56:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR120919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Dec 10 01:47:47 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	243738	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	265287	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.84	152	88258	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	112563	5.18	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	103.60%
7) Chloroethane-d5	1.93	69	125353	5.50	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	110.00%
11) 1,1-Dichloroethene-d2	2.59	63	148166	3.73	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.60%
20) 2-Butanone-d5	4.70	46	218064	58.91	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	117.82%
24) Chloroform-d	5.11	84	181122	5.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.60%
26) 1,2-Dichloroethane-d4	5.75	65	96879	5.31	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.20%
32) Benzene-d6	5.77	84	357828	5.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.60%
36) 1,2-Dichloropropane-d6	6.73	67	108459	5.36	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.20%
41) Toluene-d8	7.93	98	294147	4.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.80%
43) trans-1,3-Dichloropropene-	8.21	79	32741	3.88	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	77.60%
46) 2-Hexanone-d5	8.68	63	172292	56.42	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	112.84%
57) 1,1,2,2-Tetrachloroethane-	10.78	84	99491	5.34	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.80%
64) 1,2-Dichlorobenzene-d4	12.22	152	99360	5.85	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	117.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU120919\
Data File : VU036053.D
Acq On : 09 Dec 2019 17:38
Operator : JC/MD
Sample : K6168-11
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFE44

Quant Time: Dec 10 02:56:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR120919WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Dec 10 01:47:47 2019
Response via : Initial Calibration

