

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU120920\
 Data File : VU041518.D
 Acq On : 09 Dec 2020 16:56
 Operator : SY/MD
 Sample : L4995-01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AN7

Quant Time: Dec 10 06:55:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR120420WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Dec 10 04:45:04 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	49357	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	48870	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	25954	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	8433	3.33	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	66.60%
7) Chloroethane-d5	1.92	69	9192	4.13	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.60%
11) 1,1-Dichloroethene-d2	2.58	63	16534	2.69	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	53.80%#
20) 2-Butanone-d5	4.66	46	38336	40.33	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	80.66%
24) Chloroform-d	5.08	84	31584	4.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.20%
26) 1,2-Dichloroethane-d4	5.72	65	21620	4.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.00%
32) Benzene-d6	5.74	84	46662	4.28	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.60%
36) 1,2-Dichloropropane-d6	6.70	67	13327	4.19	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	83.80%
41) Toluene-d8	7.91	98	42850	4.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.40%
43) trans-1,3-Dichloropropene-	8.19	79	7834	4.15	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.00%
46) 2-Hexanone-d5	8.64	63	29389	37.44	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	74.88%
57) 1,1,2,2-Tetrachloroethane-	10.76	84	14498	4.13	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.60%
64) 1,2-Dichlorobenzene-d4	12.19	152	21618	4.87	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU120920\
Data File : VU041518.D
Acq On : 09 Dec 2020 16:56
Operator : SY/MD
Sample : L4995-01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AN7

Quant Time: Dec 10 06:55:37 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR120420WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Dec 10 04:45:04 2020
Response via : Initial Calibration

