

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U120920\
 Data File : VU041520.D
 Acq On : 09 Dec 2020 17:42
 Operator : SY/MD
 Sample : L4995-03
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AP4

Quant Time: Dec 10 07:01:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR120420WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Dec 10 04:45:04 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	46701	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	47497	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	21556	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	8320	3.47	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	69.40%
7) Chloroethane-d5	1.92	69	8752	4.16	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	83.20%
11) 1,1-Dichloroethene-d2	2.58	63	16339	2.81	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	56.20%#
20) 2-Butanone-d5	4.66	46	38449	42.75	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	85.50%
24) Chloroform-d	5.08	84	30811	4.65	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.00%
26) 1,2-Dichloroethane-d4	5.72	65	20762	4.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.40%
32) Benzene-d6	5.74	84	44686	4.22	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.40%
36) 1,2-Dichloropropane-d6	6.70	67	12642	4.09	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	81.80%
41) Toluene-d8	7.91	98	41019	3.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.20%
43) trans-1,3-Dichloropropene-	8.19	79	7621	4.16	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.20%
46) 2-Hexanone-d5	8.64	63	29281	38.38	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	76.76%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	13891	4.07	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	81.40%
64) 1,2-Dichlorobenzene-d4	12.19	152	18197	4.94	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU120920\
Data File : VU041520.D
Acq On : 09 Dec 2020 17:42
Operator : SY/MD
Sample : L4995-03
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AP4

Quant Time: Dec 10 07:01:15 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR120420WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Dec 10 04:45:04 2020
Response via : Initial Calibration

