

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU120920\
 Data File : VU041524.D
 Acq On : 09 Dec 2020 19:15
 Operator : SY/MD
 Sample : L4995-07
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AP8

Quant Time: Dec 10 07:13:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR120420WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Dec 10 04:45:04 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	45778	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	41312	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	21739	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	7744	3.30	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	66.00%
7) Chloroethane-d5	1.92	69	8636	4.19	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	83.80%
11) 1,1-Dichloroethene-d2	2.58	63	16418	2.88	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	57.60%#
20) 2-Butanone-d5	4.65	46	34899	39.58	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	79.16%
24) Chloroform-d	5.08	84	27266	4.20	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.00%
26) 1,2-Dichloroethane-d4	5.72	65	19912	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.40%
32) Benzene-d6	5.74	84	38619	4.20	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.00%
36) 1,2-Dichloropropane-d6	6.70	67	13099	4.87	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.40%
41) Toluene-d8	7.91	98	41018	4.55	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.00%
43) trans-1,3-Dichloropropene-	8.19	79	7666	4.81	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.20%
46) 2-Hexanone-d5	8.64	63	28039	42.26	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.52%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	11914	4.02	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	80.40%
64) 1,2-Dichlorobenzene-d4	12.19	152	18350	4.94	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU120920\
Data File : VU041524.D
Acq On : 09 Dec 2020 19:15
Operator : SY/MD
Sample : L4995-07
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AP8

Quant Time: Dec 10 07:13:15 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR120420WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Dec 10 04:45:04 2020
Response via : Initial Calibration

