

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U120920\
 Data File : VU041526.D
 Acq On : 09 Dec 2020 20:01
 Operator : SY/MD
 Sample : L4995-09
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AQ0

Quant Time: Dec 10 07:18:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR120420WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Dec 10 04:45:04 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	45474	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	43969	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	21516	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	7828	3.35	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	67.00%
7) Chloroethane-d5	1.92	69	8363	4.08	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	81.60%
11) 1,1-Dichloroethene-d2	2.58	63	16813	2.97	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	59.40%#
20) 2-Butanone-d5	4.66	46	37213	42.49	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	84.98%
24) Chloroform-d	5.08	84	29858	4.63	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.60%
26) 1,2-Dichloroethane-d4	5.72	65	20861	4.82	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.40%
32) Benzene-d6	5.74	84	43415	4.43	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.60%
36) 1,2-Dichloropropane-d6	6.70	67	11868	4.14	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	82.80%
41) Toluene-d8	7.91	98	38470	4.01	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.20%
43) trans-1,3-Dichloropropene-	8.19	79	7398	4.36	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.20%
46) 2-Hexanone-d5	8.64	63	26752	37.88	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	75.76%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	11750	3.72	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	74.40%
64) 1,2-Dichlorobenzene-d4	12.19	152	18614	5.06	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU120920\
Data File : VU041526.D
Acq On : 09 Dec 2020 20:01
Operator : SY/MD
Sample : L4995-09
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AQ0

Quant Time: Dec 10 07:18:42 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR120420WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Dec 10 04:45:04 2020
Response via : Initial Calibration

