

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120921\
 Data File : VU046218.D
 Acq On : 09 Dec 2021 14:24
 Operator : SY/MD
 Sample : M4982-07
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0CR5

Quant Time: Dec 10 03:44:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 03 05:08:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	204749	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	211513	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	111152	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	87740	52.021	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	104.040%
7) Chloroethane-d5	1.919	69	64035	49.449	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.900%
11) 1,1-Dichloroethene-d2	2.572	63	116900	38.609	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.220%
21) 2-Butanone-d5	4.639	46	129136	96.314	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	96.310%
24) Chloroform-d	5.067	84	139044	49.367	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.740%
26) 1,2-Dichloroethane-d4	5.703	65	95333	50.435	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.860%
32) Benzene-d6	5.729	84	312257	51.508	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.020%
36) 1,2-Dichloropropane-d6	6.694	67	94171	50.122	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.240%
41) Toluene-d8	7.899	98	281259	50.955	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.920%
43) trans-1,3-Dichloroprop...	8.179	79	46473	51.240	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	102.480%
47) 2-Hexanone-d5	8.632	63	90149	101.198	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	101.200%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	138940	48.763	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.520%
66) 1,2-Dichlorobenzene-d4	12.195	152	109670	51.234	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.460%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120921\
Data File : VU046218.D
Acq On : 09 Dec 2021 14:24
Operator : SY/MD
Sample : M4982-07
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0CR5

Quant Time: Dec 10 03:44:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 03 05:08:36 2021
Response via : Initial Calibration

