

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120921\
 Data File : VU046250.D
 Acq On : 10 Dec 2021 02:54
 Operator : SY/MD
 Sample : M4982-15
 Misc : 5.6mL/MSVOA_U/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 COCT0

Quant Time: Dec 10 05:26:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 03 05:08:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	159428	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	169728	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	85874	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	64456	49.080	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	98.160%
7) Chloroethane-d5	1.919	69	48535	48.134	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	96.260%
11) 1,1-Dichloroethene-d2	2.572	63	84169	35.701	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.400%
21) 2-Butanone-d5	4.639	46	100398	96.167	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	96.170%
24) Chloroform-d	5.067	84	106309	48.474	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.940%
26) 1,2-Dichloroethane-d4	5.706	65	74498	50.616	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.240%
32) Benzene-d6	5.732	84	228229	46.915	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.840%
36) 1,2-Dichloropropane-d6	6.694	67	71912	47.698	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.400%
41) Toluene-d8	7.899	98	200927	45.363	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.720%
43) trans-1,3-Dichloroprop...	8.182	79	34068	46.810	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.620%
47) 2-Hexanone-d5	8.636	63	70141	98.122	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	98.120%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	111541	48.784	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.560%
66) 1,2-Dichlorobenzene-d4	12.195	152	82116	49.654	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.300%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.584	96	4753	4.312	ug/L #	1
20) cis-1,2-Dichloroethene	4.671	96	4660	3.681	ug/L	95
25) Chloroform	5.092	83	10623	4.508	ug/L	94
30) 1,1,1-Trichloroethane	5.317	97	2831	1.359	ug/L #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120921\
Data File : VU046250.D
Acq On : 10 Dec 2021 02:54
Operator : SY/MD
Sample : M4982-15
Misc : 5.6mL/MSVOA_U/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
COCT0

Quant Time: Dec 10 05:26:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 03 05:08:36 2021
Response via : Initial Calibration

