

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120921\
 Data File : VU046209.D
 Acq On : 09 Dec 2021 10:57
 Operator : SY/MD
 Sample : VU1209MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK051

Quant Time: Dec 09 12:26:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 03 05:08:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	215881	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	221861	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.815	152	115952	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	100331	56.419	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	112.840%
7) Chloroethane-d5	1.919	69	74192	54.338	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	108.680%
11) 1,1-Dichloroethene-d2	2.575	63	132961	41.649	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.300%
21) 2-Butanone-d5	4.642	46	149055	105.438	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	105.440%
24) Chloroform-d	5.070	84	158989	53.537	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.080%
26) 1,2-Dichloroethane-d4	5.706	65	105819	53.095	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.200%
32) Benzene-d6	5.732	84	347488	54.646	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	109.300%
36) 1,2-Dichloropropane-d6	6.693	67	105714	53.642	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	107.280%
41) Toluene-d8	7.902	98	322503	55.702	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	111.400%
43) trans-1,3-Dichloroprop...	8.182	79	52652	55.345	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	110.680%
47) 2-Hexanone-d5	8.635	63	104112	111.421	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	111.420%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	156556	52.383	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	104.760%
66) 1,2-Dichlorobenzene-d4	12.195	152	124558	55.780	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	111.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120921\
Data File : VU046209.D
Acq On : 09 Dec 2021 10:57
Operator : SY/MD
Sample : VU1209MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK051

Quant Time: Dec 09 12:26:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 03 05:08:36 2021
Response via : Initial Calibration

